

T E S T R E P O R T

KOTITI No : 82211509102502-001

Issue No : 4147254241

Receipt Date : 03 / 05 / 2021

Issue Date : 10 / 05 / 2021

Sample Name	SAVEWO 3D MASK EXTREME PRO	Test Object	Research/Reference/Etc
Lot No.	LOT20240310	MDF/MFG	-
Weight/Volume	-	Expiry Date	11 / 03 / 2024
Company	The Hero Corp CO., LTD	President	Yeong ung Han
Address	8, Oegaeil-ro 21beon-gil, Mohyeon-eup, Cheoin-gu, Yongin-si, Gyeonggi-do, Republic of Korea	Telephone	82-31-329-2223

Test Item & Test Result

Test Item	Test Standards	Unit	Result	Test Method
Filtering efficiency(NaCl)	≥ 99	%	Before pretreatment : 100/99/99 After pretreatment : 99/99/99	MFDS Guideline
Filtering efficiency(Paraffin oil)	≥ 99	%	Before pretreatment : 100/100/100 After pretreatment : 100/100/100	MFDS Guideline
Facial Inhalation resistance	≤ 100	Pa	Before pretreatment : 14/13/12 After pretreatment : 13/14/13	MFDS Guideline

Researcher : Yeongmin Nam, Junhyeok Choi

Technical Manager : Man Sik Wang

- ※ The test results above shall be only subject to the requested test items.
- ※ In the case of insufficient space on the page, test items and test results may be written in the appendix.
- ※ When advertising test results or indicating test results on containers or packages, the entire content of the test report shall be indicated.
- ※ This report produced in relation to the services will not be related with MFDS. And MFDS in not legally liable.

KOTITI Testing & Research Institute



TEST REPORT

KOTITI No : 82211509102502-001

Issue No : 4147254241

Receipt Date : 03 / 05 / 2021

Issue Date : 10 / 05 / 2021

Sample photo



Researcher : Yeongmin Nam, Junhyeok Choi

Technical Manager : Man Sik Wang

- ※ The test results above shall be only subject to the requested test items.
- ※ In the case of insufficient space on the page, test items and test results may be written in the appendix.
- ※ When advertising test results or indicating test results on containers or packages, the entire content of the test report shall be indicated.
- ※ This report produced in relation to the services will not be related with MFDS. And MFDS in not legally liable.

KOTITI Testing & Research Institute

